



Amaxa® Cell Line Nucleofector® Kit L

For H9c2(2-1)


Rat heart myoblast; myoblastoid cells

Example for Nucleofection® of H9c2(2-1) cells





H9c2(2-1)cells were transfected with the Nucleofector® Kit L, Program C-020 and 2 μg of pmaxGFP® Vector. Cells were analyzed 24 hours post Nucleofection® using light (A) and fluorescence microscopy (B).



Average transfection efficiency of H9c2(2-1) cells. H9c2(2-1) cells were transfected with program C-020 and 2 µg of pmaxGFP® Vector. Cells were analyzed 24 hours post Nucleofection® by flow cytometry. Cell Viability (% PI negative) is around 90% 24 hours post Nucleofection®.

Product Description

Cat. No.		VCA-1005
Size (reactions)		25
Cell Line Nucleofector® Solution L		2.25 ml (2.05 ml + 10% overfill)
Supplement		0.5 ml (0.45 ml + 10% overfill)
pmaxGFP® Vector (0.5 µg/µl in 10 mM Tris pH 8.0)		30 µg
Certified cuvettes		25
Plastic pipettes		25
Storage and stability	Store Nucleofector® Solution, S	Supplement and pmaxGFP® Vector at 4°C. For long-term storage,

Store Nucleofector® Solution, Supplement and pmaxGFP® Vector at 4°C. For long-term storage, pmaxGFP® Vector is ideally stored at -20°C. The expiration date is printed on the solution box. Once the Nucleofector® Supplement is added to the Nucleofector® Solution it is stable for three months at 4°C.

Optimized Protocol for H9c2(2-1)

Required Material

Note

Please make sure that the entire supplement is added to the Nucleofector® Solution. The ratio of Nucleofector® Solution to supplement is 4.5:1. For a single reaction use $82\,\mu$ l of Nucleofector® Solution plus $18\,\mu$ l of supplement to make $100\,\mu$ l of total reaction volume.

- Nucleofector® Device
- Supplemented Nucleofector® Solution at room temperature
- Supplied certified cuvettes
- Supplied plastic pipettes
- Supplied pmaxGFP® Vector
- Substrate of interest, highly purified, preferably by using endotoxin-free kits; A260: A280 ratio should be at least 1.8
- 6-well culture dish or culture system of your choice
- For detaching cells: 0.5 mg/ml Trypsin and 0.2 mg/ml EDTA in PBS and supplemented culture media or PBS/0.5% BSA
- Culture medium: Dulbecco`s modified Eagle`s medium with 4 mM L-glutamine adjusted to contain 1.5 g/l sodium bicarbonate and 4.5 g/l glucose, 90%; fetal bovine serum, 10%
- Prewarm appropriate volume of culture medium to 37°C (1.5 ml per sample)
- Appropriate number of cells (4 x 10⁵ cells per sample; lower or higher cell numbers may influence transfection results)

1. Pre Nucleofection®

Cell culture recommendations

- 1.1 Replace media every 2 3 days
- 1.2 Passage cells every 2-3 days. A subcultivation ratio of 1:4-1:5 is recommended. Subculture before confluency to avoid differentiation.
- 1.3 Seed out $1.5 2 \times 10^5$ cells/T162 flask
- 1.4 Subculture 2-3 days before Nucleofection® with a ratio of 1:4-1:5

Trypsinization

- 1.5 Remove media from the cultured cells and wash cells once with PBS; use at least same volume of PBS as culture media
- 1.6 For harvesting, incubate the cells ~5 minutes at 37°C with indicated trypsinization reagent (please see required material)
- 1.7 Neutralize trypsinization reaction with supplemented culture medium or PBS/0.5% BSA once the majority of the cells (>90%) have been detached

Optimized Protocol for H9c2(2-1)

2. Nucleofection®

One Nucleofection® Sample contains

4 x 105 cells

2 μ g plasmid DNA (in 1 – 5 μ l H₂0 or TE) or 2 μ g pmaxGFP® Vector or 30 – 300nM siRNA (3 – 30 pmol/sample)

100 µl Cell Line Nucleofector® Solution L

- 2.1 Please make sure that the entire supplement is added to the Nucleofector® Solution
- 2.2 Prepare 6-well plates by filling appropriate number of wells with 1 ml of supplemented culture media and pre-incubate/equilibrate plates in a humidified 37°C/5% CO₂ incubator
- 2.3 Harvest the cells by trypsinization (please see 1.5 1.7)
- 2.4 Count an aliquot of the cells and determine cell density
- 2.5 Centrifuge the required number of cells (4×10^5) cells per sample) at 90xg for 10 minutes at room temperature. Remove supernatant completely
- 2.6 Resuspend the cell pellet carefully in 100 µl room-temperature Nucleofector® Solution per sample

Note Avoid leaving the cells in Nucleofector® Solution for extended periods of time (longer than 15 minutes), as this may reduce cell viability and gene transfer efficiency.

- 2.7 Combine 100 μ l of cell suspension with 2 μ g DNA, 2 μ g pmaxGFP® Vector or 30 nM 300 nM siRNA (3 30 pmol/sample) or other substrates
- 2.8 Transfer cell/DNA suspension into certified cuvette (sample must cover the bottom of the cuvette without air bubbles). Close the cuvette with the cap
- 2.9 Select the appropriate Nucleofector® Program C-020 (C-20 for Nucleofector® I Device)
- 2.10 Insert the cuvette with cell/DNA suspension into the Nucleofector® Cuvette Holder and apply the selected program by pressing the X-button
- 2.11 Take the cuvette out of the holder once the program is finished
- 2.12 Incubate the sample in the cuvette for 10 minutes at room temperature
- 2.13 Add \sim 500 μ l of the pre-equilibrated culture medium to the cuvette and gently transfer the sample into the prepared 6-well plate (final volume 1.5 ml media per well). Use the supplied pipettes and avoid repeated aspiration of the sample

3. Post Nucleofection®

3.1 Incubate the cells in humidified 37° C/5% CO_2 incubator until analysis. Gene expression or down regulation, respectively, is often detectable after only 4-8 hours

Additional Information

For an up-to-date list of all Nucleofector® References, please refer to: www.lonza.com/nucleofection-citations

For more technical assistance, contact our Scientific Support Team:

 USA / Canada
 Europe and Rest of World

 Phone:
 800-521-0390 (toll-free)
 Phone: +49-221-99199-400

 Fax:
 301-845-8338
 Fax: +49-221-99199-499

Lonza Cologne AG 50829 Cologne, Germany

 $Please \ note that \ the \ Amaxa^{\otimes} \ Nucleo fector^{\otimes} \ Technology \ is \ not intended \ to \ be used for \ diagnostic purposes or for testing \ or \ treatment \ in \ humans.$

The Nucleofector® Technology, comprising Nucleofection® Process, Nucleofector® Device, Nucleofector® Solutions, Nucleofector® 96-well Shuttle® System and 96-well Nucleocuvette® plates and modules is covered by patent and/or patent-pending rights owned by Lonza Cologne AG.

Amaxa, Nucleofector, Nucleofection and maxGFP are either registered trademarks or trademarks of the Lonza Cologne AG in Germany and/or U.S. and/or other countries.

Other product and company names mentioned herein are the trademarks of their respective owners.

This kit contains a proprietary nucleic acid coding for a proprietary copepod fluorescent protein intended to be used as a positive control with this Lonza product only. Any use of the proprietary nucleic acid or protein other than as a positive control with this Lonza product is strictly prohibited. USE IN ANY OTHER APPLICATION REQUIRES A LICENSE FROM EVROGEN. To obtain such a license, please contact Evrogen at license@evrogen.com.

The CMV promoter is covered under U.S. Patents 5,168,062 and 5,385,839 and its use is permitted for research purposes only. Any other use of the CMV promoter requires a license from the University of Iowa Research Foundation, 214 Technology Innovation Center, Iowa City, IA 52242.

The use of this product in conjunction with materials or methods of third parties may require a license by a third party. User shall be fully responsible for determining whether and from which third party it requires such license and for the obtainment of such license.

No statement is intended or should be construed as a recommendation to infringe any existing patent.

© Copyright 2009, Lonza Cologne AG. All rights reserved DCL-1015 05/09